

# Scottish Health Technical Memorandum 2010

(Part 1 of 6)

Overview and management responsibilities

# **Sterilization**

#### Disclaimer

The contents of this document are provided by way of guidance only. Any party making any use thereof or placing any reliance thereon shall do so only upon exercise of that party's own judgement as to the adequacy of the contents in the particular circumstances of its use and application. No warranty is given as to the accuracy of the contents and the Property and Environment Forum Executive, which produced this document on behalf of NHSScotland Property and Environment Forum, will have no responsibility for any errors in or omissions therefrom.

The production of this document was jointly funded by the Scottish Executive Health Department and the NHSScotland Property and Environment Forum.



# **Contents**

1.	Steril	ization and the role of management	page 4
1.1	Introd	uction	
1.4	The E	uropean Union Directives on medical devices	
1.6	Defini	tion of medical device	
1.8	The th	ree Directives	
1.9	The re	egulatory framework	
1.13	Impac	t on sterilization	
1.14	Summ	nary of management responsibilities	
2.	Steril	izers – an overview	page 8
2.1	Introd	uction	
2.7	Clinica	al sterilizers using high-temperature steam	
2.10		s loads	
2.13	Fluids		
2.16		apped instruments and utensils	
2.20		al sterilizers using hot air	
2.24		al sterilizers using low-temperature steam and for	maldehyde
2.28		al sterilizers using ethylene oxide	
2.35		atory sterilizers/autoclaves	
2.38	•	ating cycles	
2.42		re media preparator	
2.45		steamer	
2.46	Anıma	al house sterilizer	
3.		tory requirements	page 16
3.1	Introd		
3.2		and safety	
	3.5	Health and Safety at Work etc Act 1974	
	3.6	Management of Health and Safety at Work Reg 1999	gulations
	3.8	Workplace (Health, Safety and Welfare) Regula	ations 1992
	3.10	Provision and Use of Work Equipment Regulation	ons 1998
	3.12	Pressure Systems Safety Regulations 2000	
	3.14	Control of Substances Hazardous to Health Re 1999	gulations
	3.17	Reporting of Injuries, Diseases and Dangerous Regulations 1995	Occurrences
	3.20	Manual Handling Operations Regulations 1992	
	3.22	Personal Protective Equipment at Work Regula	tions 1992
3.23	Medic	inal products	
	3.23	Medicines Act 1968	

page 30



**References** 

	3.24	Medicines (Standard Provisions of Licences and Certificates) Amendment (No 3) Regulations 197	7
	3.25	Medicines (Standard Provision of Licences and C Amendment Regulations 1992	
3.26	Consu	umer protection	
	3.27	Consumer Protection Act 1987	
3.29	Electr	omagnetic Compatibility Regulations 1992	
3.32	Active	Implantable Medical Devices Regulations 1992	
4.	Britis	h and European standards	page 22
4.1	Introd	uction	
4.4	Europ	ean standards	
5	Perso	onnel	page 23
5.1	Introd	uction	
5.2	Traini	ng	
5.5	Functi	ional responsibility	
5.11	Key p	ersonnel	
	5.12	Management	
	5.13	User	
	5.16	Competent person (pressure vessels)	
	5.22	Authorised person (sterilizers)	
	5.27	Test person (sterilizers)	
	5.30	Maintenance person (sterilizers)	
	5.34	Microbiologist (sterilizers)	
	5.37	Personnel for medicinal products	
	5.38	Production manager	
	5.39	Quality controller	
	5.43	Other personnel	

Version 2.0: June 2001 Page 3 of 42



# 1. Sterilization and the role of management

#### Introduction

"The fundamental cause of this disaster is to be found in human failings ranging from simple carelessness to poor management of men and plant. The Committee heard of no imminent technological advance in the field of production of intravenous fluids which will eliminate the need for skilful men devoted to their work ... Too many people believe that sterilization of fluids is easily achieved with simple plant operated by men of little skill under a minimum of supervision ... Public safety in this, as in many other technological fields, depends ultimately on untiring vigilance ..."

- 1.1 The quotation above comes from the principal conclusions of the committee chaired by Sir Cecil Clothier and appointed to investigate an incident in which five patients died as a result of a faulty sterilizer. The tragedy led to a thorough overhaul of the methods of managing sterilizers, among which was the revision of Health Technical Memorandum 2010 (then HTM 10), the last edition of which was published in 1980.
- 1.2 No disaster on a comparable scale has been reported since. Nonetheless, both the law and public opinion are now less forgiving of lapses than they were two decades ago. Tighter statutory control, resulting from new European Union (EU) Directives, will soon extend to almost every aspect of sterilization, and practices which were common a few years ago will no longer be acceptable or even lawful.
- 1.3 The science and art of sterilization are complex and subtle. The testing, maintenance and reporting procedures described in this SHTM may seem excessive to some, but they are based upon good practice in both the UK and Europe, as formalised in European Standards designed to support the new EU Directives.

# The European Union Directives on medical devices

- 1.4 Until now, statutory controls on the practice of sterilization, other than in the manufacture of medical products, have been few. The major Acts and Regulations which are likely to affect the management of a sterilizer are described in Chapter 3, but specific references to sterilization in the legislation were rare. This has changed as a series of three EU Directives came into effect regulating the safety, quality and effectiveness of medical devices.
- 1.5 This section summarises basic information about the Directives. Further details are available from the Medical Devices Agency of the Department of Health, England, which has a UK remit.



#### **Definition of medical device**

- 1.6 The Directives define a medical device as any instrument, apparatus, appliance, material or other article, whether used alone or in combination, including the software necessary for its proper application intended by the manufacturer to be used on human beings for the purpose of:
  - a. diagnosis, prevention, monitoring, treatment or alleviation of disease;
  - b. diagnosis, monitoring, treatment, alleviation or compensation of an injury or handicap;
  - c. investigation, replacement or modification of the anatomy or of a physiological process;
  - d. control of conception;

and which does not achieve its principal intended action in or on the human body by pharmacological, immunological or metabolic means, but which may be assisted in its function by such means (Council Directive 93/42/EEC).

1.7 The Directives apply equally to "accessories". An accessory is defined as "an article which, whilst not being a device, is intended specifically by its manufacturers to be used together with a device to enable it to be used in accordance with the use of the device intended by the manufacturer of the device".

#### The three Directives

- 1.8 The three EU directives are as follows:
  - a. the Active Implantable Medical Devices Directive (Council Directive 90/385/EEC) covers all powered implants or partial implants that are left in the human body. (Heart pacemakers are the most common example of powered implants.) The directive was adopted by the EU council on 20 June 1990 and came into effect in the UK on 1 January 1993 as the Active Implantable Devices Regulations 1992 (see paragraph 3.32);
  - the Medical Devices Directive (Council Directive 93/42/EEC) covers most other medical devices ranging from first aid bandages and tongue depressors through to hip prostheses and will therefore have a wide impact on sterilization. The directive was adopted by the EU council on 14 June 1993. It came into effect on 1 January 1995;
  - c. the In Vitro Diagnostic Medical Devices Regulations 2000 (SI 1315) implements the council directive 98/79/EC and covers any medical device, reagent product, kit, instrument, apparatus or system which is intended to be used in vitro for the examination of substances derived from the human body. Some examples of in-vitro diagnostic devices are blood group reagents, pregnancy test kits, and hepatitis B test kits.



# The regulatory framework

- 1.9 The Directives set out the essential requirements that devices must not compromise the health or safety of the patient, user or any other person and that any risks associated with the device are compatible with patient health and protection. Any side-effects must be acceptable when weighed against the intended performance.
- 1.10 Devices meeting these requirements will be entitled to carry the "CE" marking, signifying that the device satisfies the requirements essential for it to be fit for its intended purpose. All devices except custom-made devices and devices intended for clinical trials, ("investigations" in the directive), whether used in public-sector or private-sector hospitals and nursing homes, or sold in retail outlets, will have to carry the "CE" marking.
- 1.11 Adoption of the Directives has meant that the UK's voluntary system of manufacturer registration and product approval for controlling certain medical devices used by the NHS has been replaced by a more comprehensive statutory system covering all devices used in the UK. The Medical Devices Agency (MDA) of the Department of Health, is the competent authority to carry out the requirements of the Directives in the UK. The main role of the MDA is to ensure compliance with the UK regulations, evaluate vigilance reports received from manufacturers, and carry out a preclinical assessment of devices intended for clinical investigation. The MDA is also responsible for approving the independent certification organisations (the notified bodies) that will check and prove that defined classes of medical devices meet the essential requirements and thus enable manufacturers to apply the "CE" marking to their products.
- 1.12 The Medical Devices Directive includes a classification system whereby the level of regulatory control applied to devices is proportional to the degree of risk inherent in the device. The strictest controls will therefore only apply to the limited number of high-risk products.

# Impact on sterilization

1.13 Managers who ensure that their machines and procedures comply with the guidance in this SHTM should have no difficulty in complying with the Medical Devices Directive.



# Summary of management responsibilities

- 1.14 SHTM 2010 will assist managers and other personnel to ensure that sterilizers are operated safely and effectively and in compliance with existing and anticipated legislation and standards. To this end, the major responsibilities of management can be summarised as follows:
  - a. to ensure that sterilization is carried out in compliance with the law and policy of the Scottish Executive Health Department;
  - to ensure that all personnel connected with sterilization, whether NHS employees or contract personnel, are suitably qualified and trained for their responsibilities;
  - to ensure that purchased sterilizers conform to legal requirements, the minimum specifications set out in British and European standards, and any additional requirements of the Scottish Executive Health Department
  - d. to ensure that sterilizers are installed correctly and safely with regard to proper functioning, safety of personnel and environmental protection;
  - e. to ensure that newly installed sterilizers are subject to a documented scheme of validation comprising installation checks and tests, commissioning tests and performance qualification tests before they are put into service;
  - f. to ensure that sterilizers are subject to a documented scheme of periodic tests at yearly, quarterly, weekly and (in some cases) daily intervals:
  - g. to ensure that sterilizers are subject to a documented scheme of preventative maintenance;
  - to ensure that procedures for production, quality control and safe working are documented and adhered to in the light of statutory requirements and accepted best practice;
  - to ensure that procedures for dealing with malfunctions, accidents and dangerous occurrences are documented and adhered to.



#### 2. Sterilizers - an overview

#### Introduction

- 2.1 This Scottish Health Technical Memorandum groups sterilizers into two broad categories according to their use:
  - clinical sterilizers are designed to process medical devices, medicinal products and other goods and materials that are used in the clinical care of patients;
  - b. **laboratory sterilizers** are designed to process goods and materials and are not directly used in the clinical care of patients.
- 2.2 Their operation should be kept strictly separate. Loads intended for processing in a clinical sterilizer should not be put into a laboratory sterilizer, and vice-versa.
- 2.3 Sterilizers can also be classified according to the sterilizing agent (the sterilant) used:
  - a. high-temperature steam;
  - b. low-temperature steam and formaldehyde;
  - c. ethylene oxide.
- 2.4 High-temperature steam is the sterilant of choice because of its superior performance. Machines using other sterilants should be reserved either for loads which would be damaged by exposure to high-temperature steam (such as certain surgical devices) or for loads that would not be sterilized by exposure to high-temperature steam (such as certain non-aqueous fluids).
- 2.5 Clinical sterilizers are available employing any one of the four sterilants. The laboratory sterilizers described in this SHTM use only high-temperature steam.
- 2.6 Guidance on selection and specification, operational management, validation and verification is given in the other parts of this SHTM.



# Clinical sterilizers using high-temperature steam

- 2.7 These are by far the most common sterilizers used in the NHS, and are manufactured in three basic types according to the nature of load they are designed to process: porous loads, fluids, or unwrapped instruments and utensils. The operating cycles are designed to cope with the differing properties of the various types of load. It is essential that a sterilizer is used only for the type of load for which it is designed.
- 2.8 High-temperature steam inactivates pathogens by a combination of moisture and heat. The process is well understood and the attainment of sterilization conditions can normally be confirmed by simple physical measurements. (This is not so for sterilizers using chemical sterilants, where microbiological test procedures are necessary.)
- 2.9 High-temperature steam sterilizers are large machines requiring permanently installed engineering services (including good-quality steam) and purpose-built accommodation. Some smaller models are transportable and generate steam from an internal reservoir.

#### **Porous loads**

- 2.10 Clinical sterilizers using high-temperature steam to process porous loads are commonly known as "porous load sterilizers". They are intended to deal with porous items such as towels, gowns and dressings; and medical and surgical equipment, instruments and utensils packaged or wrapped in porous materials such as paper or fabrics.
- 2.11 Sterilization is achieved by direct contact of the load items with good-quality saturated steam at a preferred sterilization temperature of 134°C.
- 2.12 As porous loads trap both air and moisture, an efficient and reliable air removal system is essential. An air detector is fitted to ensure that the operating cycle does not proceed until sufficient air and other non-condensable gases have been removed from the chamber and load. The correct functioning of the air detector is crucial to the performance of the sterilizer.

#### **Fluids**

- 2.13 Clinical sterilizers using high-temperature steam to process aqueous fluids are commonly known as "fluids sterilizers". They are used to sterilize fluids in sealed containers (normally bottles) of either glass or plastic. They operated at a preferred sterilization temperature of 121°C.
- 2.14 Fluids in glass containers can be hazardous. At a temperature of 121°C the pressure inside a one-litre bottle having a normal fill of fluid is approximately 4 bar. If the door were to be opened at this temperature, and the load



exposed to ambient air, the thermal stresses arising in the glass would be sufficient to crack the bottle and cause an explosion. A temperature of 80°C is regarded as a safe maximum at which the door can be opened (even at this temperature the pressure inside a one-litre bottle is still 1.8 bar). Fluid sterilizers are fitted with a thermal door-lock to ensure that when glass containers are being processed the door cannot be opened until the temperature inside all the containers has fallen below 80°C. Failure to observe this requirement has led to serious accidents resulting from the explosion of glass containers.

2.15 Fluids in plastic containers present less of a hazard. Operating cycles for plastic containers allow the door to be opened when the temperature inside the containers falls below 90°C.

#### Unwrapped instruments and utensils

- 2.16 This type of sterilizer is used to process unwrapped surgical components intended for immediate use. Sterilization is achieved by the direct contact of the component with saturated steam at a preferred sterilization temperature of 134°C.
- 2.17 These sterilizers should not be used to process wrapped instruments and utensils, where the wrapping could inhibit the removal of air and the penetration of steam. Neither should they be used for unwrapped instruments and utensils with narrow lumens, where air removal and steam penetration would similarly be impaired.
- 2.18 Since the sterilized instruments and utensils are exposed to the air on being removed from the chamber, they are susceptible to immediate recontamination. These sterilizers are therefore suitable for clinical use only within the immediate environment in which the instruments are to be used. Wherever possible, instruments and utensils should be wrapped and processed in a porous load sterilizer.
- 2.19 Transportable (bench-top) models are electrically heated, requiring only a 13 Amp socket-outlet and no piped services. They are commonly used in theatre suites where there is no central supply service and in primary healthcare units such as general practitioners' and dentists' surgeries.

# Clinical sterilizers using hot air

2.20 Clinical sterilizers using hot air as a sterilant are correctly known as "dry-heat sterilizers", and sometimes as "hot-air sterilizers" or "sterilizing ovens". They are intended to process materials such as oils, powders and some ophthalmic instruments, which can withstand high temperatures but are likely to be damaged or not sterilized by contact with steam. They operate at a preferred sterilization temperature of 160°C.



- 2.21 They are not suitable for use as drying cabinets (see BS 2648 for specifications for drying cabinets).
- 2.22 Dry-heat sterilizers are essentially electric ovens and are therefore simpler than the other pressure sterilizers described in this SHTM. A filter and fan are used to maintain the chamber slightly above atmospheric pressure to ensure that the sterility of the product and the integrity of the clean-room environment are not compromised. Although the cycle is under automatic control, the operator is allowed considerable freedom in selecting the required combination of sterilization temperature and time. Recommended combinations are shown in Table 2.1 and advice on their selection is given in Part 4 of this SHTM.
- 2.23 Dry-heat sterilizers are not efficient. It is difficult to obtain an even temperature distribution within the chamber, air circulation is inhibited when the chamber is full (even with a circulating fan), and heat transfer from the air to the load can be very slow. A complete cycle, including cooling to 80°C, takes approximately eight hours for a full test load as described in Part 3 of this SHTM. If this time is unacceptable, a sterilizer fitted with assisted cooling is recommended, reducing the cycle time for the same load to approximately five hours.

# Clinical sterilizers using low-temperature steam and formaldehyde

- 2.24 Heat-sensitive materials (wrapped or unwrapped) which will withstand saturated steam at temperatures up to 80°C are normally processed in either low-temperature steam disinfectors ("LTS disinfectors") or low-temperature steam and formaldehyde sterilizers ("LTSF sterilizers"). Sterilizers designed for LTSF will normally incorporate an LTS disinfection cycle.
- 2.25 Disinfection is achieved by the direct contact of the load with saturated steam at a minimum temperature of 71°C at sub-atmospheric pressure. Sterilization is achieved by contact with both saturated steam and formaldehyde gas. Either process may also be used to decontaminate soiled surgical components before they are washed and reprocessed.
- 2.26 Formaldehyde is a toxic gas. Part 5 of this SHTM contains safety information.
- 2.27 Since the sterilization process is ultimately dependent on chemical action, microbiological test methods are required to confirm that sterilization conditions have been attained.

**NOTE:** Despite their name, LTSF sterilizers are disinfectors.



# Clinical sterilizers using ethylene oxide

- 2.28 Clinical sterilizers using ethylene oxide gas as a sterilant are commonly known as "ethylene oxide sterilizers" or "EO sterilizers".
- 2.29 EO sterilizers are used to process heat-sensitive materials and devices which cannot withstand low-temperature steam. They should not be used to process items which can be sterilized by alternative methods, that is, by high-temperature steam, dry heat or LTSF. They should not be used to resterilize items which have been sterilized by irradiation.
- 2.30 EO sterilizers are used extensively in industrial manufacture of sterile medical devices but are relatively uncommon in hospitals. Two classes of EO sterilizers are suitable for NHS use:
  - a. small sterilizers, of chamber volumes around 150 litres, where the sterilant is pure EO at sub-atmospheric pressure supplied from a disposable cartridge contained within the chamber;
  - b. large sterilizers, of chamber volume up to 500 litres, where the sterilant is either pure EO or EO diluted with another gas, supplied from cylinders. EO sterilizers have the potential to cause serious environmental pollution. Sterilizers using chlorofluorocarbon (CFC) gases as diluents should no longer be installed.
- 2.31 EO is a highly reactive liquid and gas which is toxic, flammable and explosive. The safe operation of EO sterilizers requires careful consideration of all aspects of the installation and operation of equipment.
- 2.32 The entire EO process is complex and requires specialised facilities for washing, packaging and preconditioning loads before processing and degassing before use. Large sterilizers will also require additional plant to dispose safely of exhaust products.
- 2.33 The efficacy of the process is affected by the packaging used to wrap goods for sterilization. Since the sterilization process is ultimately dependent upon chemical action, microbiological test methods are required to confirm that sterilization conditions have been attained.
- 2.34 Managers considering installing EO sterilizers should be aware of the following points:
  - a. the difficulty in validating and monitoring suitable cleaning processes for loads before they are sterilized;
  - b. the difficulty in carrying out representative performance qualification tests for the wide variety of loading conditions that may be used;
  - c. the difficulty in carrying out meaningful bioburden studies on small numbers of widely differing devices to be sterilized;



 the problems associated with determining the levels of residual EO and its reaction products when small numbers of widely differing devices are processed.

#### Laboratory sterilizers/autoclaves

- 2.35 Laboratory sterilizers, also known as autoclaves, are used for making-safe discard material and processing apparatus and materials to be used within clinical laboratories. They are not intended for the sterilization of medical devices or medicinal products intended for the clinical care of patients.
- 2.36 Unlike clinical sterilizers, the laboratory sterilizers covered in this SHTM are designed for use only with high-temperature steam. No chemical sterilants are used.
- 2.37 Certain common laboratory operations may be carried out more economically with specialised machines designed for the purpose, and these are described below.

### **Operating cycles**

- 2.38 Laboratory sterilizers are often required to process a wide range of materials and objects, and they are equipped with one or more operating cycles each designed for a particular application. Different types of load generally require different operating cycles. Cycles are normally preset, and proceed automatically once selected and started.
- 2.39 The range of cycles that a sterilizer can provide will depend on details of its construction. For example, the methods used to remove air from the chamber, the means employed to cool and dry the load, and the provision of safety features.
- 2.40 Laboratory sterilizers may be equipped with one or more of the following operating cycles:
  - a. make-safe of small plastic discard;
  - b. make-safe of contained fluid discard:
  - c. sterilization of culture media;
  - d. disinfection of fabrics;
  - e. sterilization of glassware and equipment;
  - f. free steaming.
- 2.41 Guidance on the specification of operating cycles is given in Part 2 of this SHTM.



# **Culture media preparator**

- 2.42 Many of the problems which relate to sterilizing culture media can be solved by the use of small sterilizers in which the media constituents are placed directly into the chamber, thus avoiding the use of glass containers and their attendant hazards.
- 2.43 The machine consists of two or three modules incorporated into a system designed to provide controlled preparation, sterilization, cooling and dispensing of culture media with a minimum of attention by the operator. The system may also include a module which automatically stacks the completed culture plates.
- 2.44 The sterilizer module is essentially a pressure-cooker in which water and dehydrated culture media are mixed, sterilized and then cooled to below 80°C. This type of sterilizer is particularly suitable for manufacturing batches of culture media in volumes between 1 and 20 litres.

#### Köch steamer

2.45 A Köch steamer is designed to expose a load to steam at near-atmospheric pressure and is commonly used for melting solidified agar. Steamers are not sterilizers and the product cannot be regarded as sterile. No further information specific to Köch steamers is given in this SHTM.

#### Animal house sterilizer

2.46 The very wide range of materials and implements used in the care of laboratory animals is often catered for by specialised sterilizers with capacities as high as 10 m³, which run several operating cycles. Examples of loads include bedding for discard, fresh bedding, feed bottles, food and water, cages, and tools and implements for use by personnel in the animal house. In view of the specialised nature of these machines, no further information specific to animal house sterilizers is given in this SHTM. Users are advised to adapt the guidance on laboratory sterilizers to their circumstances in consultation with the authorised person.



#### **Table 2.1 Sterilization temperature bands**

	High-temperature steam		Dry heat			LTS	LTSF	Ethylene oxide		
Sterilization temperature (°C) <sup>a</sup>	115	121	126	134	160	170	180	71 <sup>b</sup>	71	30-56
Maximum allowable temperature (°C)	118	124	129	137 <sup>c</sup>	170	180	190	80	80	d
Minimum holding time (min)	30	15	10	3	120	60	30	10	180 <sup>e</sup>	f

#### Notes:

- a. The temperature setting on the automatic controller will not generally be the sterilization temperature, but a higher temperature within the sterilization temperature band.
- b. Disinfection temperature.
- c. British Standards permit 138°C.
- d. For EO, the maximum allowable temperature will normally be 4°C above the sterilization temperature.
- e. For LTSF, the sterilization conditions may specify either a continuous holding time or the number of pulses for formaldehyde required to achieve sterilization.
- f. For EO, the "gas exposure time" is determined for each sterilizer by microbiological methods during commissioning but is typically 2-7 hours depending upon sterilization temperature and gas concentration.



# 3. Statutory requirements

#### Introduction

3.1 So far as sterilization is concerned, the chief areas of legislation with which managers should be familiar are health and safety, medicinal products and consumer protection.

#### Health and safety

- 3.2 The largest body of law with which managers need to be familiar concerns health and safety, in particular the Health and Safety at Work etc Act 1974 (the HSW Act) and its various regulations.
- 3.3 The HSW Act and its regulations require employers to assess the risk to their employees. Attention is drawn to the following hazards which are implicit in the practice of sterilization:
  - a. the hazard of scalding from escaping steam;
  - b. the high temperatures (up to 200°C) at which sterilizers are operated;
  - c. the stored energy hazards associated with the operation of pressure vessels contained within all steam and some EO sterilizers;
  - d. the stored energy hazards associated with the pressurised containers in which EO gas is transported;
  - e. the explosive hazards associated with the sterilization of fluids in sealed glass bottles;
  - f. the toxic properties of formaldehyde gas used in low-temperature steam and formaldehyde (LTSF) sterilizers;
  - g. the toxic and explosive properties of ethylene oxide gas used in ethylene oxide (EO) sterilizers;
  - h. the infection hazard associated with the microbial pathogens that may be handled by personnel using certain laboratory sterilizers;
  - the hazard of infection to patients and staff by the inadvertent release of an unsterile load due to the failure of a sterilization and quality control process;
  - j. the hazards associated with the handling of heavy and hot loads while loading and unloading sterilizers.
- 3.4 The guidance given throughout this SHTM is designed to ensure that these hazards are minimised and that sterilization procedures comply with the relevant legislation and established good practice.



#### Health and Safety at Work etc Act 1974

The HSW Act sets out the basic legal responsibilities of employers and employees with regard to health and safety at work.

#### Management of Health and Safety at Work Regulations 1999

- 3.6 The Management of Health and Safety at Work Regulations 1999 (SI 3242) expand upon the principles of the HSW Act.
- 3.7 The core of the regulations is a requirement of employers to make a systematic assessment of the risks to health and safety of their employees and others, arising from work activities.

#### Workplace (Health, Safety and Welfare) Regulations 1992

- 3.8 The Workplace (Health, Safety and Welfare) Regulations 1992 (SI 1992/3004) aim to ensure that workplaces meet the health, safety and welfare needs of each member of the workforce, including people with disabilities.
- 3.9 Most of the regulations deal with the physical requirements of the workplace. Managers concerned with the operation of sterilizers should pay particular attention to the regulations and maintenance, ventilation, temperature, lighting, cleanliness, room dimensions and space, floors, doors and traffic routes.

#### **Provision and Use of Work Equipment Regulations 1998**

- 3.10 The Provision and Use of Work Equipment Regulations 1998 (PUWER) aim to ensure the provision of safe work equipment and its safe use.
- 3.11 PUWER 98 replaces the existing Provision and Use of Work Equipment Regulations 1992 and applies to all equipment (including lifting equipment) used at work in the health care sector. PUWER 92's requirements are carried forward in full but there are important new additions, including a requirement to inspect work equipment where significant risk could result from incorrect installation or relocation; deterioration; or as a result of exceptional circumstances; and to record the results those inspections (Regulation 6).

#### **Pressure Systems Safety Regulations 2000**

3.12 The regulations on pressure systems apply to all steam sterilizers, to EO sterilizers operating above 0.5 bar, and to the steam and compressed air services. They replace the Pressure Systems and Transportable Gas Containers Regulations 1989, which had earlier replaced the sections of the Factories Act 1961 that were relevant to steam sterilizers. Transportable gas containers were also removed from the scope of the 1989 Regulations and are now covered by the Carriage of Dangerous Goods (Classification,



Packaging and Labelling) and Use of Transportable Pressure Receptacles Regulations 1996. This would apply to cartridges and cylinders used to supply sterilant or purging gas to EO sterilizers.

3.13 The regulations also define the duties of the competent person: a competent individual person (other than an employee) or a competent body of persons corporate or unincorporate.

#### **Control of Substances Hazardous to Health Regulations 1999**

- 3.14 Schedule 1 of the Control of Substances Hazardous to Health (COSHH)
  Regulations lists ethylene oxide and formaldehyde as two substances
  hazardous to health which are subject to a maximum exposure limit for
  inhalation. These limits are reviewed annually and updated by amendments
  to the regulations. The current limits (1999) are given in Table 3.1. These
  limits must not be regarded as safe work exposures.
- 3.15 The Health and Safety Executive (HSE) publishes an annually updated guidance note on current exposure limits, *Occupational exposure limits (EH 40)*.
- 3.16 Users of laboratory sterilizers should note that a "substance hazardous to health" may include a micro-organism which creates a hazard to the health of any person. Guidance on the precautions to be taken when handling micro-organisms may be found in the Health and Safety Committee (HSC) documents, Categorisation of pathogens according to hazard and categories of containment, (second edition 1990) complied by the Advisory Committee on Dangerous Pathogens, and Safe working and the prevention of infection in clinical laboratories, compiled by the Health Services Advisory Committee.

# Reporting of Injuries, Diseases and Dangerous Occurrences Regulations 1995

- 3.17 Commonly known as RIDDOR, these regulations impose duties on persons responsible for the activities of persons at work, and on self-employed persons, to report accidents resulting in death or major injury arising out of or in connection with work, and to report specified dangerous occurrences. They also require certain particulars of accidents at work to be reported both to the Department of Health and also to the Health and Safety Executive, and require records to be kept.
- 3.18 Steam and certain EO sterilizers contain pressure vessels as defined under Part 1 of Schedule 1.
- 3.19 Poisoning by ethylene oxide is a reportable disease listed under Schedule 2.



#### **Manual Handling Operations Regulations 1992**

- 3.20 The regulations require employers to make an ergonomic assessment of all manual handling operations which involve a risk injury, and to reduce the risk as far as is reasonably practicable. Factors to be assessed include the nature of the task, the load, the working environment and individual capability.
- 3.21 Managers should assess the risks associated with loading and unloading sterilizers, whether by loading trolleys or by hand. Top-loading sterilizers can be especially hazardous if lifting equipment is not available. The mass of the load is not the only source of risk; the temperature and other factors should be taken into account. Risks associated with maintenance and overhauling should also be assessed. Reference should also be made to the Lifting Operations and Lifting Equipment Regulations 1998 (LOLER).

#### Personal Protective Equipment at Work Regulations 1992

3.22 Managers should assess whether the risks associated with sterilization require the use of personal protective equipment (PPE). Some examples include heat-resistant gloves for use when hot loads are removed from sterilizers, protective gloves for use when handling discard material in laboratories, eye or face protection when testing sterilizers containing fluids in glass bottles, and foot protection of operators loading and unloading sterilizers.

# **Medicinal products**

#### **Medicines Act 1968**

3.23 Where a sterilizer is to be used to sterilize medicinal products, the licensing provisions of the Medicines Act 1968 apply. Further information may be found in, *Guidance to the NHS on the licensing requirements of the Medicines Act 1968*, published by the Medicines Control Agency.

# Medicines (Standard Provisions of Licences and Certificates) Amendment (No 3) Regulations 1977

3.24 The Medicines (Standard Provisions of Licences and Certificates)
Amendments (No 3) Regulations 1977 introduced a qualified person who, in certain circumstances, has statutory responsibility for quality control in the manufacture of medicinal products (see Chapter 5). This will include decisions on release of a sterilized product.



#### Medicines (Standard Provision of Licences and Certificates) Amendment Regulations 1992

3.25 The Medicines (Standard Provisions of Licences and Certificates)
Amendment Regulations 1992 (SI 1992/2846) give statutory force to the commission document, *The rules governing medicinal products in the European Community Volume IV: Guide to good manufacturing practice for medicinal products*. All provisions in the guide came into force on or before 1 January 1993. The annex on sterilization contains requirements that are implemented by the guidance in this SHTM.

# **Consumer protection**

3.26 In recent years, new legislation has been introduced affording protection to persons who may be harmed by unsafe goods supplied to them. In certain circumstances this may include products from sterilizers.

#### **Consumer Protection Act 1987**

- 3.27 Part 1 implements EU Council Directive 85/374/EEC (the Product Liability Directive) providing for compensation to be paid to persons injured by a defective product. Under the Act a product is defective "if the safety of the product is not such as persons generally are entitled to expect", taking the circumstances into account. It is likely that civil action for damages could be taken against a hospital for supplying, for example, "sterile" products that were not in fact sterile and caused the infection of a patient.
- 3.28 Part 2 introduces a "general safety requirement" on the suppliers of "consumer goods" only. It is a criminal offence to supply unsafe consumer goods, whether or not actual harm has been caused. Consumer goods are defined as "any goods which are ordinarily intended for private use or consumption", and are regarded as unsafe when "they are not reasonably safe having regard to all the circumstances". It is not clear whether products from hospital sterilizers are to be regarded as consumer goods. (Controlled drugs and licensed medicinal products are exempt from Part 2 since they are governed by other legislation.)

# **Electromagnetic Compatibility Regulations 1992**

- 3.29 The Electromagnetic Compatibility Regulations (SI 2372) (the EMC Regulations), impose requirements concerning the electromagnetic compatibility of most types of electrical and electronic apparatus which must be complied with, before such apparatus is to be supplied or taken into service.
- 3.30 A sterilizer (and any ancillary equipment) is a "relevant apparatus" within the terms of the regulations, and will have to meet standards of emission of an immunity to electromagnetic disturbance. Note that it is an offence not only



to supply but also to "take into service" a sterilizer that does not conform to the regulations.

3.31 The regulations do not apply to any sterilizer supplied to be taken into service in the EU before 28 October 1992. A sterilizer supplied or taken into service in the UK on or before 31 December 1995 is not required to comply with the regulations provided it complies with the requirements of the Wireless Telegraphy Acts listed in Schedule 1 of the regulations.

**NOTE:** Detailed guidance on the application of the EMC regulations in healthcare premises may be found in SHTM 2014; *Abatement of electrical interference*.

#### **Active Implantable Medical Devices Regulations 1992**

- 3.32 The Active Implantable Medical Devices Regulations 1992 (SI 3146) set out the essential requirements which active implantable medical devices (such as heart pacemakers) must satisfy before they can be placed on the market or put into service.
- 3.33 Schedule 2, paragraph 7 requires such devices to be designed, manufactured and packed in a non-reusable packaging according to procedures which are sufficient to ensure that:
  - a. the device is sterile when placed on the market; and
  - b. if handled in accordance with conditions as to storage and transport laid down by the manufacturer, the device remains sterile until the packaging is removed and the device is implanted.
- 3.34 Schedule 2, paragraph 14 sets out requirements for the labelling of sterile packs.

Table 3.1 Maximum exposure limits at atmospheric formaldehyde and ethylene oxide

Gas	Short-term e	Short-term exposure limits		ng-term exposure limits	
	[ppm]	[mg m <sup>-3</sup> ]	[ppm]	[mg m <sup>-3</sup> ]	
Formaldehyde	2	2.5	2	2.5	
Ethylene oxide	_	_	5	9.2	

#### Notes:

The short-term exposure limit (STEL) is the average exposure over any 15-minute period.

The long-term exposure limit (STEL) is the exposure over any 24-hour period expressed as a single uniform exposure over an 8-hour period.

COSHH does not specify a STEL for EO. In such cases the STEL is deemed to be three times the LTEL in accordance with the recommendations of the Health and Safety Executive.

Source: HSE guidance note EH40 (Feb 1999).



# 4. British and European standards

#### Introduction

- 4.1 Industry standards for sterilization have developed rapidly since the publication of HTM 10 in 1980. British standards which existed at that time have been thoroughly revised and extended. European standards now cover not only design, construction, performance and safety, but also validation, routine testing and operation.
- 4.2 British and European standards, supplemented by specific requirements for the NHS, form the basis of the guidance given in the 'Design considerations' part of this SHTM.
- 4.3 The main standards for sterilizers are BS 3970 for clinical sterilizers and BS 2646 for laboratory sterilizers.

# **European standards**

- 4.4 European standards on sterilization are generally more extensive than British standards in specifying not only design, construction, performance and safety requirements of sterilizers, but also that persons responsible for sterilization operate a quality system and that part of that system is validation and routine testing of the process.
- 4.5 While the guidance given here is designed to conform broadly with European standards, SHTM 2010 must not be regarded as a substitute for the standards themselves.



#### 5. Personnel

#### Introduction

5.1 This chapter introduces the personnel who may share the responsibility for the safe and efficient operation of sterilizers. It gives guidance on qualifications and training and summarises areas of responsibility.

# **Training**

- 5.2 It is essential that personnel at all levels have a sound general knowledge of the principles, design and functions of sterilizers. They should be trained on those types and models of sterilizers with which they are concerned. They should have some knowledge of the basic elements of microbiology in order to ensure personal safety, safety of others and general safety. Training given to individuals should be recorded and reviewed regularly.
- Accredited courses on sterilization, suitable for personnel at all levels, are run by various training providers. Further information is available from the NHS in Scotland Property and Environment Forum Executive and the authorised persons (sterilizers).
- 5.4 Detailed training on particular models of sterilizer is usually available from the manufacturer, either on-site (such as during validation) or by courses at their premises.

# **Functional responsibility**

- There have been profound changes in the management philosophy of the NHS over recent years, and there is a trend towards deregulation and contracting-out of services. It is not possible to prescribe a management structure for sterilization that is universally applicable given the wide range of circumstances in which a sterilizer may be employed, from a busy sterile services department in a major general hospital to a small rural dental practice.
- 5.6 The approach chosen for this SHTM is to identify the distinct functions that need to be exercised and the responsibilities that go with them. The titles given are therefore generic; they describe the individual's role in connection with sterilization, but are not intended to be prescriptive job titles for terms of employment, indeed, many of the personnel referred to may not be resident staff but employed by outside bodies and working on contract. Some of them will have other responsibilities unconnected with sterilization and in some cases the same individual may take on more than one role.



- In every case, however, it is possible to identify a **user** who is responsible for the day-to-day management of the sterilizer. The philosophy of this SHTM is to invest the user with the responsibility for seeing that the sterilizer is operated safely and efficiently.
- 5.8 The law requires that a **competent person (pressure vessels)** who is not the **user** is designated to exercise certain responsibilities of inspection for all steam sterilizers and other sterilizers containing pressure vessels.
- 5.9 For small installations where the user is qualified to perform all required test and maintenance functions, no other personnel may be necessary. This may be satisfactory for small sterilizers run by dentists or general practitioners. However, it is strongly recommended that in all cases the user receive professional advice from an **authorised person (sterilizers)**, and that testing and maintenance be carried out by a suitably qualified **test person (sterilizers)** and a **maintenance person (sterilizers)** with assistance from a **microbiologist (sterilizers)** where microbiological testing is required.
- 5.10 Where the sterilizer is used to manufacture medicinal products, the functions of the user are exercised by a **production manager** and a **quality controller**.

### **Key personnel**

5.11 For the purposes of SHTM 2010, the following are the key roles in the management of sterilization.

#### Management

5.12 Management is defined as the owner, occupier, employer, general manager, chief executive or other person who is ultimately accountable for the sole operation of its premises.

#### User

- 5.13 The user is defined as the person designated by management to be responsible for the sterilizer.
- In a hospital, the user could be a sterile services department manager, laboratory manager or theatre manager; in primary care he or she could be a general practitioner, dentist, or other health professional. Where a sterilizer is used to process medicinal products, the user is normally the production manager in charge of the entire manufacturing process.



- 5.15 The principal responsibilities of the user are as follows:
  - a. to certify that the sterilizer is fit for use;
  - b. to hold all documentation relating to the sterilizer, including the names of other key personnel;
  - c. to ensure that the sterilizer is subject to periodic testing and maintenance:
  - d. to appoint operators where required and ensure that they are adequately trained:
  - e. to maintain production records;
  - f. to establish procedures for product release (for medical products, in cooperation with the quality controller).

#### **Competent person (pressure vessels)**

- 5.16 The competent person (pressure vessels) is defined as a person or organisation designated by the management to exercise certain legal responsibilities with regard to the written scheme of examination of any pressure vessel associated with a sterilizer described in the Pressure Systems Safety Regulations 2000. The shorter term "competent person" is used in this SHTM.
- 5.17 The competent person should not be the user, nor any of the other key personnel associated with the sterilizer in question.
- 5.18 The HSC Approved Code of Practice, 'Safety of pressure systems' requires the competent person to have the necessary expertise, knowledge and experience for the system in question.
- 5.19 The principal duties of the competent person under the regulations are as follows (they need not all be exercised by the same individual):
  - a. advising on the scope of the written scheme of examination;
  - b. drawing up the written scheme of examination or certifying the scheme as being suitable;
  - c. carrying out examinations in accordance with the written scheme, assessing the results and reviewing the written scheme for its suitability.
- 5.20 Most insurance companies maintain a technical division able to advise on appointing a competent person. The authorised person (sterilizers) will also be able to provide advice.
- 5.21 Further information about the written scheme of examination will be found in Part 4 of this SHTM.



### **Authorised person (sterilizers)**

- The authorised person (sterilizers) is defined as a person designated by management to provide independent auditing and advice on sterilizers and sterilization and to review and witness documentation on validation. The shorter term "authorised person" is used in this SHTM.
- 5.23 The authorised person should:
  - a. have a minimum of two years recent experience in the validation of sterilization processes to modern standards;
  - b. have a degree in a relevant science subject or corporate membership of a relevant professional institution;
  - c. have completed an accredited course for authorised persons (sterilizers) and successfully passed the examination;

#### or alternatively, should:

- d. have applied for registration as an authorised person (sterilizers) no later than 31 December 1994;
- e. have at least ten years experience in the validation of porous load and laboratory sterilization processes;
- f. have two years experience in a responsible position:
- g. successfully pass an accredited examination for authorised persons (sterilizers) within five years of registration.
- The authorised person is required to liaise closely with other professionals in various disciplines and consequently, the appointment should be made known in writing to all interested parties. He or she should have direct contact with the user and other key personnel.
- 5.25 The principal responsibilities of the authorised person are as follows:
  - a. to provide general and impartial advice on all matters concerned with sterilization;
  - b. to advise on programmes of validation;
  - c. to audit reports on validation, revalidation and yearly tests prepared by the test person;
  - d. to advise on programmes of periodic tests and periodic maintenance;
  - e. to advise on operational procedures for routine production.
- 5.26 A register of suitably qualified authorised persons is maintained by the Institute of Healthcare Engineering and Estate Management (IHEEM).



#### **Test person (sterilizers)**

- 5.27 The test person (sterilizers) is defined as a person designated by management to carry out validation and periodic testing of sterilizers. The shorter term "test person" is used in this SHTM.
- 5.28 The test person should:
  - a. be qualified to at least HNC in engineering or microbiological sciences;
  - b. have completed an accredited course for test persons (sterilizers) and successfully passed the examination;
  - c. have been recently employed in an NHS hospital with responsibility for validation and periodic testing for one or more sterilization processes;

#### or alternatively:

- d. have a certificate demonstrating satisfactory completion of an accredited course (City and Guilds or equivalent) in the validation and periodic testing of at least two sterilization processes;
- have at least three years experience in the validation and periodic testing of porous load sterilizers and at least one other sterilization process.
- 5.29 The principal responsibilities of the test person are as follows:
  - a. to conduct the validation tests specified in Part 3 of this SHTM and to prepare the validation report;
  - b. to conduct the periodic tests specified in Part 3 and to prepare reports as required by the user;
  - c. to conduct any additional tests at the request of the user.

#### **Maintenance person (sterilizers)**

- 5.30 The maintenance person (sterilizers) is defined as a person designated by management to carry out maintenance duties on sterilizers. The shorter term "maintenance person" is used in this SHTM.
- 5.31 The maintenance person should be a fitter or an electrician with documentary evidence to demonstrate competence in the maintenance of one or more types of sterilizer. He or she should be in a position to deal with any breakdown in an emergency and have the ability to diagnose faults and carry out repairs or to arrange for repairs to be carried out by others.
- 5.32 The principal responsibilities of the maintenance person are as follows:
  - a. to carry out the maintenance tasks outlined in Part 4;
  - b. to carry out additional maintenance and repair work at the request of the user.



A maintenance person who has a minimum of two years experience in the maintenance of sterilizers and who has obtained a recognised qualification in the testing of sterilizers may perform the duties of the test person for the daily, weekly and quarterly tests described in Part 3.

#### **Microbiologist (sterilizers)**

- The microbiologist (sterilizers) is defined as a person designated by management to be responsible for advising the user on microbiological aspects of the sterilization of non-medical products. The shorter term "microbiologist" is used in this SHTM.
- 5.35 The microbiologist should have a relevant degree (for example microbiology or medicine) and will normally be a member of the hospital staff.
- 5.36 The principal responsibilities of the microbiologist are as follows:
  - a. to advise the user on the microbiological aspects of sterilization procedures for non-medicinal products;
  - to arrange for the culturing of biological indicators used in microbiological tests (normally low-temperature steam and formaldehyde (LTSF) and ethylene oxide (EO) sterilizers);
  - c. to audit the documentation from all sterilizers which have been tested by microbiological methods.

#### Personnel for medicinal products

5.37 Where a sterilizer is to be used in the productron of medicinal products, the provisions of the Medicines Act 1968 apply. The responsibilities that would otherwise be exercised by the user are divided between the production manager and the quality controller. Guidance on the duties of each can be found in the EU commission document, *Guide to good manufacturing practice for medicinal products*.

#### **Production manager**

5.38 The production manager is defined as a person designated by management to be responsible for the production of medicinal products.

#### **Quality controller**

- 5.39 The quality controller is defined as a person designated by management to be responsible for quality control and medicinal products with authority to establish, verify and implement all quality control and quality assurance procedures.
- 5.40 He or she should have the authority, independent of the production manager, to approve materials and products and to reject, as seen fit, raw materials, packaging materials, and intermediate, bulk and finished products



not complying with the relevant specification or not manufactured in accordance with approved procedures.

- 5.41 The quality controller should be professionally qualified (for example in pharmacy). Any additional qualifications will depend on the type of licence which is held, for example:
  - a. where a product licence is held, the quality controller should satisfy the requirements of the qualified person as defined in the Medicines (Standard Provisions of Licences and Certificates) Amendment (No. 3) Regulations 1977. If the quality controller does not meet these requirements, a qualified person should be designated to exercise the functions specified in the regulations;
  - b. where the manufacturer's licence "specials" is held, as is generally the case in hospitals, the quality controller need not satisfy the requirements of a qualified person.
- Further information about qualified person can be found in MAL 45 Medicines Acts 1968, 1971.

#### Other personnel

- 5.43 The following personnel are also mentioned in this SHTM.
- The **laboratory safety officer** is defined as a person designated by management to be responsible for all aspects of laboratory safety including equipment, personnel and training relating to safety issues, and ensuring compliance with safety legislation and guidelines.
- 5.45 An **operator** is defined as any person with the authority to operate a sterilizer, including the noting of sterilizer instrument readings and simple housekeeping duties.
- 5.46 The **manufacturer** is defined as a person or organisation responsible for the manufacturer of a sterilizer.
- 5.47 The **contractor** is defined as a person or organisation designated by management to be responsible for the supply and installation of the sterilizer, and for the conduct of the installation checks and tests. The contractor is commonly the manufacturer of the sterilizer.



# References

#### NOTE:

Where there is a requirement to address a listed reference, care should be taken to ensure that all amendments following the date of issue are included.

Publication ID	Title	Publisher	Date	Notes			
Acts and Regulations							
	Building (Scotland) Act	HMSO	1959				
	Clean Air Act	HMSO	1993				
	Consumer Protection Act	HMSO	1987				
	Electricity Act	HMSO	1989				
	Health and Safety at Work etc Act	HMSO	1974				
	Health and Medicines Act	HMSO	1988				
	Registered Establishments (Scotland) Act	HMSO	1998				
	Water (Scotland) Act	HMSO	1980				
SI 3146	Active Implantable Medical Devices Regulations	HMSO	1992				
SI 1995	Active Implantable Medical Devices (Amendment and Transitional Provisions) Regulations 1995	HMSO	1995				
SI 2179 & 187	Building Standards (Scotland) Regulations (as amended)	HMSO	1990				
SI 2092	Carriage of Dangerous Goods (Classification, Packaging & Labelling) and Use of Transportable Pressure Receptacles Regulations	HMSO	1996				
	Building Standards (Scotland) Regulations: Technical Standards Guidance	HMSO	1998				
SI 1460	Chemicals (Hazard Information and Packaging for Supply) Regulations (CHIP2)	HMSO	1997				
SI 3140	Construction (Design and Management) Regulations	HMSO	1994				
SI 437	Control of Substances Hazardous to Health Regulations (COSHH)	HMSO	1999				
SI 635	Electricity at Work Regulations	HMSO	1989				
SI 1057	Electricity Supply Regulations (as amended)	HMSO	1988				



Publication ID	Title	Publisher	Date	Notes
SI 2372	Electromagnetic Compatibility Regulations (as amended)	HMSO	1992	
SI 2451	Gas Safety (Installation and Use) Regulations	HMSO	1998	
SI 917	Health & Safety (First Aid) Regulations	HMSO	1981	
SI 682	Health & Safety (Information for Employees) Regulations	HMSO	1989	
SI 1380	Health and Safety (Training for Employment) Regulations	HMSO	1990	
SI 341	Health and Safety (Safety Signs and Signals) Regulations	HMSO	1996	
SI 2792	Health and Safety (Display Screen Equipment) Regulations	HMSO	1992	
SI 1315	In Vitro Diagnostic Medical Devices Regulations 2000	HMSO	2000	
SI 3232	Ionising Radiations Regulations 1999	HMSO	1999	
SI 2307	Lifting Operations and Lifting Equipment Regulations (LOLER)	HMSO	1998	
SI 2051	Management of Health and Safety at Work Regulations	HMSO	1999	
SI 2793	Manual Handling Operations Regulations	HMSO	1992	
SI 3017	Medical Devices Regulations	HMSO	1994	
SI 1790	Noise at Work Regulations	HMSO	1989	
SI 2966	Personal Protective Equipment at Work (PPE) Regulations	HMSO	1992	
SI 2306	Provision and Use of Work Equipment Regulations (PUWER)	HMSO	1998	
SI 3139	Personal Protective Equipment (EC Directive) Regulations	HMSO	1992	
SI 128	Pressure Systems Safety Regulations (PSSR)	HMSO	2000	
SI 3163	Reporting of Injuries, Diseases and Dangerous Occurrences Regulations (RIDDOR)	HMSO	1995	
SI 119	Water Supply (Water Quality) (Scotland) Regulations	HMSO	1990	
SI 3004	Workplace (Health, Safety and Welfare) Regulation	HMSO	1992	



Publication ID	Title	Publisher	Date	Notes				
British Standa	British Standards							
BS 593	Specification for laboratory thermometers		1989					
BS 1781	Specification for linen and linen union textiles		1981					
BS 2646	Autoclaves for sterilization in laboratories Part 1: Specification for design, construction, safety and performance	BSI Standards	1993					
	Part 2: Guide to planning and installation		1990					
	Part 3: Guide to safe use and operation		1993					
	Part 4: Guide to maintenance		1991					
	Part 5: Methods of testing for function and performance		1993					
BS 2648	Performance requirements for electrically heated laboratory drying ovens (PD2517,6/56)	BSI Standards	1955					
BS 2775	Specification for rubber stoppers and tubing for general laboratory use		1987					
BS 3693	Recommendations for design of scales and indexes on analogue indicating instruments		1992					
BS 3928	Method for sodium flame test for air filters (other than for air supply to I.C. engines and compressors)	BSI Standards	1969					



Publication ID	Title	Publisher	Date	Notes
BS 3970	Sterilizing and disinfecting equipment for medical products Part 1: Specification for general requirements	BSI Standards	1990	
	Part 2: Specification for steam sterilizers for aqueous fluids in sealed rigid containers		1991	
	Part 3: Specification for steam sterilizers for wrapped goods and porous loads		1990	
	Part 4: Specification for transportable steam sterilizers for unwrapped instruments and utensils		1990	
	Part 5: Specification for low temperature steam disinfectors		1993	
	Part 6: Specification for sterilizers using low temperature steam with formaldehyde			
BS 4196-0	Sound power level of noise sources. Guide for the use of basic standards and for the preparation of noise test codes	BSI Standards	1981	
BS 4275	Guide to implementing an effective respiratory protective device programme	BSI Standards	1997	
BS 5164	Specification for indirect acting electrical indicating and recording instruments and their accessories		1975	
BS 5295	Environmental cleanliness in enclosed spaces			
	Part 1: Specification for clean rooms and clean air devices		1989	
BS 5304	British standard code of practice for safety of machinery	BSI Standards	1988	
BS 5815	Sheets, sheeting, pillowslips, towels, napkins and continental quilts secondary covers  Parts 1: Specification for sheeting	BSI Standards	1989 1988	
	etc		1991	
	Part 2: specification for towels etc.			
	Part 3: Specification for counterpanes etc.			



Publication ID	Title	Publisher	Date	Notes
BS 6000	Guide for the selection of an acceptance sampling system, scheme or plan for inspection of discrete items in lots	BSI Standards	1996	
BS 6001	Sampling procedures for inspection by attributes	BSI Standards	1991	
BS 6068	Water quality Sect.1.2 Glossary	BSI Standards	1997	
	Sect 6.5 Guidance on sampling of drinking water and water used for food processing		1991	
	<b>Sect. 6.7</b> Guidance on sampling of water and steam in boiler plants.		1994	
BS 6257	Specification for paper bags for steam sterilization for medical use		1989	
BS 6447	Specification for absolute and gauge pressure transmitters with electrical outputs		1984	
BS 7671	Requirements for electrical installations. IEE wiring regulations	BSI Standards	1992	16 <sup>th</sup> edition
BS 7720	Specification for non-biological sterilization indicators equivalent to the Bowie and Dick Test		1995	
BS EN 134	Respiratory protective devices. Nomenclature of components. Names of components in three CEN languages and diagrams for respiratory protective equipment	BSI Standards	1998	
BS EN 285	Sterilization, steam sterilizers, large sterilizers	BSI Standards	1997	
BS EN 550	Sterilization of medical devices. Validation and routine control of sterilization by ethylene oxide	BSI Standards	1994	
BS EN 552	Sterilization of medical devices. Validation and routine control of sterilization by irradiation	BSI Standards	1994	
BS EN 554	Sterilization of medical devices. Validation and routine control of sterilization by moist heat	BSI Standards	1994	



Publication ID	Title	Publisher	Date	Notes
BS EN 556	Sterilization of medical devices. Requirements for terminally sterilized medical devices to be labelled 'STERILE'	BSI Standards	1995	
BS EN 724	Guidance on the application of EN 29001 and EN 46001 and of EN 29002 and EN 46002 for non-active medical devices	BSI Standards	1995	
BS EN 764	Pressure equipment. Terminology and symbols: pressure, temperature, volume	BSI Standards	1995	
BS EN 837-1	Bourdon tube pressure gauges: dimensions, metrology, requirements and testing	BSI Standards	1998	
BS EN 866	Biological systems for testing sterilizers and sterilization processes  Part 1: General requirements	BSI Standards	1997	
	Part 2: Particular systems for use in ethylene oxide sterilizers		1998	
	Part 3: Particular systems for use in moist heat sterilizers		1997	
BS EN 867	Non-biological systems for use in sterilizers Part 1: General requirements	BSI Standards	1997	
	Part 2: Process indicators			
	Part 3: Specification for Class B indicators for use in the Bowie and Dick test			
BS EN 868	Packaging materials and systems for medical devices which are to be sterilized. General requirements	BSI Standards	1997	
BS EN 980	Graphical symbols for the use in the labelling of medical devices	BSI Standards	1997	
BS EN 1174	Sterilization of medical devices. Estimation of population of microorganisms on product	BSI Standards	1996	
BS EN 1422	Sterilizers for medical purposes – ethylene oxide sterilizers – specification	BSI Standards	1998	
BS EN 22872	Complete, filled transport packages.  Method for determination of resistance to compression	BSI Standards	1993	



Publication ID	Title	Publisher	Date	Notes
BS EN 25667-1	Water quality. Guidance on design of sampling programmes	BSI Standards	1994	
BS EN 25667-2	Water sampling . Guidance on sampling techniques	BSI Standards	1993	
BS EN 30993	Biological evaluation of medical devices Part 3: Tests for genotoxicity, carcinotoxicity, and reproductive toxicity	BSI Standards	1994	
	Part 4: Selection of tests for interaction with blood		1994	
	Part 5: Tests for cytotoxicity, in vitro methods		1994	
	Part 6: Tests for local effects after implantation		1995	
BS EN ISO 3746	Acoustics. Determination of sound power levels of noise sources using sound pressure. Survey method using an enveloping measurement surface over a reflecting plane	BSI Standards	1996	
BS EN 45003	Calibration and testing laboratory accreditation systems, general requirements for operation and recognition	BSI Standards	1995	
BS EN 45011	General requirements for bodies operating product certification systems	BSI Standards	1998	
BS EN 45012	General requirements for bodies operating assessment and certification/registration of quality system	BSI Standards	1998	
BS EN 45014	General criteria for supplier's declaration of conformity	BSI Standards	1993	
BS EN 45020	Standardization and related activities	BSI Standards	1998	
BS EN 46001	Specification for the application of EN ISO9001 to the manufacture of medical devices	BSI Standards	1997	
BS EN 46002	Specification for the application of EN ISO9002 to the manufacture of medical devices	BSI Standards	1994	



Publication ID	Title	Publisher	Date	Notes
BS EN 50081	Electromagnetic compatibility. Generic emission standard Part 1: Residential, commercial and light industry	BSI Standards	1992	
	Part 2: Industrial environment		1994	
BS EN 50082	Electromagnetic compatibility. Generic immunity standard Part 1: Residential, commercial and light industry	BSI Standards	1998	
	Part 2: Industrial environment		1995	
BS EN 60079-14	Electrical apparatus for explosive gas atmospheres. Electrical installations in hazardous areas (other than mines)	BSI Standards	1997	
BS EN 60581-2	Thermocouples. Manufacturing tolerances	BSI Standards	1996	
BS EN 60584-1	Thermocouples reference table	BSI Standards	1996	
BS EN 60651	Specification for sound level meters	BSI Standards	1994	
BS EN 60751	Industrial platinum resistance thermometer sensors		1996	
BS EN 60804	Specification for integrating averaging sound level meters		1994	
BS EN 61010	Safety requirements for electrical equipment for measurement, control and laboratory use			
	-1: General requirements		1993	
	<ul> <li>-2-041: Particular requirements for autoclaves and sterilizers using steam for the treatment of medical materials and for laboratory processes</li> </ul>		1997	
	<ul> <li>-2-042: Particular requirements for autoclaves and sterilizers using toxic gas for the treatment of medical materials and for laboratory processes</li> </ul>		1997	
	<ul> <li>-2-043: Particular requirements for autoclaves and sterilizers using either hot air or hot inert gas for the treatment of medical materials and for laboratory processes</li> </ul>		1998	



Publication ID	Title	Publisher	Date	Notes
BS EN ISO 9001	Quality systems. Model for quality assurance in design, development, production, installation and servicing	BSI Standards	1994	
BS EN ISO 9002	Quality systems. Model for quality assurance in production, installation and servicing	BSI Standards	1994	
European Uni	on Directives			
65/65/EEC	Approximation of provisions laid down by law, regulation or administrative action relating to proprietary medicinal products.	Official Journ European Communities (OJEC), no 2 9/2/65, p 369	; 22,	
75/107/EEC	Approximation of the laws of member states relating to bottles used as measuring containers.	Official Journ European Communities (OJEC), L42,	;	
90/385/EEC	Approximation of the laws of the Member States relating to active implantable medical devices.	Official Journ European Communities (OJEC), L189 20/7/90, p 17	s 9	
91/356/EEC	Laying down the principles and guidelines of good manufacturing practice for medicinal products for human use.	Official Journ European Communities (OJEC). L193 17/7/91, p 30	; 3	
80/778/EEC	Quality of water intended for human consumption	Official Journ European Communities		
93/42/EEC	Medical Devices Directive	Official Journ European Communities (OJEC), L169 12/7/93, p 1	<b>;</b>	
98/79/EC	In Vitro Diagnostic Medical Devices Directive	Official Journ European Communities (OJEC), L33 <sup>-1</sup> 7/12/98	s,	
Scottish Health Technical Guidance				
SHTM 2007	Electrical services supply and distribution	P&EFEx	2001	CD-ROM
SHTM 2011	Emergency electrical services	P&EFEx	2001	CD-ROM
SHTM 2014	Abatement of electrical interference	P&EFEx	2001	CD-ROM



Publication ID	Title	Publisher	Date	Notes
SHTM 2020	Electrical safety code for low voltage systems (Escode – LV)	P&EFEx	2001	CD-ROM
SHTM 2022	Medical gas pipeline systems	P&EFEx	2001	CD-ROM
SHTM 2023	Access and accommodation for engineering services	P&EFEx	2001	CD-ROM
SHTM 2025	Ventilation in healthcare premises	P&EFEx	2001	CD-ROM
SHTM 2030	Washer-disinfectors	P&EFEx	2001	CD-ROM
SHTM 2027	Hot and cold water supply, storage and mains services	P&EFEx	2001	CD-ROM
SHTM 2030	Washer-disinfectors	P&EFEx	2001	CD-ROM
SHTM 2031	Clean steam for sterilizers	P&EFEx	2001	CD-ROM
SHTM 2040	Control of legionellae in healthcare premises – a code of practice	P&EFEx	2001	CD-ROM
SHTM 2045	Acoustics	P&EFEx	2001	CD-ROM
SHPN 1	Health service building in Scotland	HMSO	1991	
SHPN 2	Hospital briefing and operational policy	HMSO	1993	
SHPN 13	Sterile services department	HMSO	1994	
SHPN 15	Accommodation for pathology services	HMSO	1994	
SHTN 1	Post commissioning documentation for health buildings in Scotland	HMSO	1993	
SHTN 2	Domestic hot and cold water systems for Scottish Health Care Premises	EEF	1999	CD-ROM
SHTN 4	General Purposes Estates and Facilities Model Safety Permit-to-Work system	EEF	1998	CD-ROM
SHGN	'Safe' hot water and surface temperatures	EEF	1999	CD-ROM
	NHS in Scotland – PROCODE	P&EFEx	2001	Version 1.1
NHS in Scotla	nd Firecode			
SHTM 81	Fire precautions in new hospitals	P&EFEx	1999	CD-ROM
SHTM 82	Alarm and detection systems	P&EFEx	1999	CD-ROM
SHTM 83	Fire safety in healthcare premises: general fire precautions	P&EFEx	1999	CD-ROM
SHTM 84	Fire safety in NHS residential care properties	P&EFEx	1999	CD-ROM
SHTM 85	Fire precautions in existing hospitals	P&EFEx	1999	CD-ROM
SHTM 86	Fire risk assessment in hospitals	P&EFEx	1999	CD-ROM
SHTM 87	Textiles and furniture	P&EFEx	1999	CD-ROM



Publication ID	Title	Publisher	Date	Notes
SFPN 3	Escape bed lifts	P&EFEx	1999	CD-ROM
SFPN 4	Hospital main kitchens	P&EFEx	1999	CD-ROM
SFPN 5	Commercial enterprises on hospital premises	P&EFEx	1999	CD-ROM
SFPN 6	Arson prevention and control in NHS healthcare premises	P&EFEx	1999	CD-ROM
SFPN 7	Fire precautions in patient hotels	P&EFEx	1999	CD-ROM
SFPN 10	Laboratories on hospital premises	P&EFEx	1999	CD-ROM
UK Health Tee	chnical Guidance			
HBN 29	Accommodation for pharmaceutical services	HMSO	1988	
HTM 67	Building components: laboratory fitting out system	HMSO	1993	
CONCODE	Contracts and commissions for the NHS estate – contract procedures	HMSO	1994	
MES	Model Engineering Specifications	NHS Estates	1997	As required
MES C14	Sterilizers	NHS Estates	1993	
HBN 13	Supplement 1 – Ethylene oxide sterilization section	HMSO	1994	
MES C02	Thermal insulation	NHS Estates	1993	
Health and Sa	fety Guidance			
EH 40	HSE Occupational Exposure limits	HSE	Annual	
HN(76)126	Hospital design note 4 (noise control): amendments to appendices II, IV and VII	DHSS	1976	
STB3A/85/12	Performance and safety specification for media sterilizers. Media devices directorate	DHSS	1985	
	Emmerson, A. M. Sterilization, disinfection and cleaning of medical equipment: guidance on decontamination from the Microbiology Committee to the Department of Health Medical Devices Directorate. Medical devices directorate	Department of Health	1993	
	Biological tests for graded milk. Memo 139/Foods.	Ministry of Health	1937	



Publication ID	Title	Publisher	Date	Notes
	Scottish Infection Manual Guidance on the core standards for the control of infection in hospitals, healthcare premises and at the community interface	The Scottish Office	1998	
	Programmable electronic systems in safety related applications: General technical guidelines	HSE	1987	
	Programmable electronic systems in safety related applications: an introductory guide	HSE	1987	
L 5	General COSHH ACOP (Control of substances hazardous to health) Carcinogens ACOP (Control of carcinogenic substances) and Biological agents ACOP (Control of biological agents) Control of Substances Hazardous to Health Regulations1999 Approved Code of Practice	HSE	1999	
L 22	Safe use of work equipment: Approved code of practice and guidance	HSE	1998	
L 23	Manual handling operations: guidance on regulations	HSE	1998	
L 24	Workplace health, safety and welfare: Approved code of practice and guidance	HSE	1992	
L25	Personal protective equipment at work at work: guidance on regulations		1992	
L113	Safe use of lifting equipment: Approved code of practice and guidance	HSE	1998	
L122	Safety of pressure systems: Pressure Systems Safety Regulations 2000. Approved Code of Practice	HSE Books	2000	
PM73	Safety at Autoclaves	HSE Books	1998	
	Precautions for work with human and animal. Transmissible Spongiform Encaphalopathies	HSE (ACDP)		
	Categorisation of pathogens according to hazard and categories of containment	HSE (ACDP)	1995	4 <sup>th</sup> Edition
	Safe working and the prevention of infection in clinical laboratories	HSC (HSAC)		
HS(R)23	A guide to the 'Reporting of Injuries, Diseases and Dangerous Occurrences Regulations 1995	HSE	1995	



Publication ID	Title	Publisher	Date	Notes
Miscellaneou	s References	1	•	•
GGMP Volume IV	Guide to good manufacturing practice for medicinal products – The rules governing medicinal products in the European Community			
	Atomic absorption spectrophotometry 1979 version	HMSO	1979	(out of print)
	Cadmium in potable waters by atomic absorption spectrophotometry 1976	HMSO	1976	(out of print)
	Colour and turbidity of waters 1981	HMSO	1981	(out of print)
	Determination of anions and cations, transition metals, and other complex ions and organic acids and bases in water by chromatography 1990	HMSO	1990	
	Lead in potable waters by atomic absorption spectrophotometry 1976	HMSO	1976	(out of print)
	Lead and cadmium in fresh waters by atomic absorption spectrophotometry (second edition) a general introduction to electrothermal atomization atomic absorption spectrophotometry 1986	HMSO	1986	(out of print)
	Measurements of electrical conductivity and the laboratory determination of the pH value of natural, treated and waste waters.	HMSO		(out of print)
	Mercury in waters, effluents, soils and sediments etc, additional methods	HMSO	1985	(out of print)
	Phosphorus and silicon in waters, effluents and sludges 1992	HMSO	1993	
	Model Water Byelaws: Dept. of the Environment	HMSO	1986	
LG 2	Lighting guide: hospitals and healthcare buildings	Chartered Institution of Building Services Engineers	1989	
	Sterilization and disinfection of heat-liable equipment	Central Sterilizing Club	1986	