Incident Reporting and Investigation Centre (IRIC)



# Safety Information Message

Reference: SIM2303 Issued: 01 August 2023 Review Date: 01 August 2024

# OPSS Product Recall for Vevor-branded products: risk of electric shock, fire, burns and injuries

## **Summary**

The Office of Product Safety and Standards has initiated a Product Recall for Vevor-branded products presenting a risk of electric shock, fire, burns and injuries. OPSS has highlighted the possible use of Vevor-branded ultrasonic cleaners within the NHS organisation.

# **Action**

- 1. Bring to the attention of the relevant managers.
- 2. Refer to the OPSS product recall page for <u>Vevor-branded</u> products and follow the required actions.
- 3. The ultrasonic cleaners detailed in the product recall are outlined in the equipment details.

# **Equipment details**

Vevor-branded ultrasonic cleaners

PSD Case Number	Product Name	Model Number	Link to Product Recall Page
2203-0390	Vevor Ultrasonic Cleaner	JPS-10A	<u>Link</u>
2208-0327	Vevor Digital Ultrasonic Cleaner	(JM) 020 Barcodes: 220311006,	<u>Link</u>
		220601540021	
2208-0328	Vevor Ultrasonic Cleaner	JP-040S	<u>Link</u>
2210-0002	Vevor Digital Ultrasonic Cleaner	JPS-100A	Link
2211-0013	Vevor Ultrasonic Jewellery Cleaner	JPS-30A	Link
2211-0024	Vevor Digital Ultrasonic Cleaner	TH-20A LS	Link
2211-0027	Vevor Digital Ultrasonic Cleaner	TH-SPQXJ-30A	Link
2211-0046	Vevor Ultrasonic Cleaner	JPS-20A	Link
2212-0026 / 2208-0327	Vevor Ultrasonic Cleaner	(JM)-020	Link

# **Background**

The Office for Product Safety and Standards (OPSS) have identified Vevor-branded products as presenting a risk to consumer safety. These products have been recalled by the manufacturer and/or additional businesses involved in the importation, supply and distribution of affected products.

FAC406-275, v4 Page **1** of **2** 

Reference: SIM2303 Issued: 01 August 2023 Review Date: 01 August 2024

An OPSS Vevor-branded Product Recall page has been produced for each product. They provide further information on the product identifiers, the risks presented by the product and the corrective action to be taken.

The OPSS highlighted specific concerns of the use of Vevor-branded ultrasonic cleaners within the NHS organisation affecting the safety of staff.

## **Suggested onward distribution**

Health & Safety Medical Physics Safety Representatives

Laboratories Risk Management

#### **Enquiries**

Enquiries and adverse incident reports should be addressed to:

#### Incident Reporting & Investigation Centre (IRIC)

NHS National Services Scotland

Tel: 0131 275 7575 Email: <a href="mailto:nss.iric@nhs.scot">nss.iric@nhs.scot</a>

**Accessibility**: Please contact us using the above details if you are blind or have a sight impairment and would like to request this alert in a more suitable format.

**IRIC remit**: general information about adverse incidents, safety alerts and IRIC's role can be found in <u>CEL 43 (2009)</u>, Safety of Health, Social Care, Estates and Facilities Equipment: NHS Board and Local Authority Responsibilities. issued 30 October 2009.

Report an incident: Information on how to report an adverse incident

NHS National Services Scotland is the common name for the Common Services Agency for the Scotlish Health Service <a href="https://www.nss.nhs.scot/">https://www.nss.nhs.scot/</a>

FAC406-275, v4 Page 2 of 2